

Table 1: Cardiac Implantable Electronic Devices—Associated Complications

<p>Procedure related</p>	<p>Reaction to medications, anesthesia</p> <p>Bleeding/hematoma formation</p> <p>Pneumothorax/hemothorax</p> <p>Pericardial effusion/tamponade</p> <p>Diaphragmatic stimulation</p> <p>Cardiac arrhythmias</p>
<p>Device related</p>	<p>Lead perforation</p> <p>Lead infection</p> <p>Lead displacement/dislodgement</p> <p>Inappropriate function</p>
<p>Miscellaneous</p>	<p>Twiddler’s syndrome</p> <p>Tricuspid regurgitation</p> <p>Electromagnetic interference</p>

